

50 TD GOLD GLUCOSE CONTROL SOLUTION LEVEL 2 SET- glucose dehydrogenase, glucose
EPS Bio Technology Corp.

Home Use Medical Device - 50 TD Gold Glucose Control Solution Level 2

50 TD Gold Glucose Control Solution Level 2 Box



TD Gold Glucose Control Solution Level 2 label

SIZE: 55 X 30 mm

TD Gold

Glucose Control Solutions

Level 2 4ml

Recap the bottle immediately after use.
Discard test strips in 90 days after first opening date.

Open on /

LOT

EXP

- TD Gold Glucose Control Solution is for in vitro diagnostic use only.
- Rock the control solution bottle back and forth several times before use.
- DO NOT freeze.
- Use the control solution only for testing with TD Gold Meters and Test Strips.
- Store the TD Gold Glucose Control Solution between 36-86°F (2-30°C).

Manufactured for Total Diabetes, LLC
Telephone Assistance / 855-525-8332

Manufacturer / EPS Bio Technology Corp.
PN 706400019_1249_02

50 TD GOLD GLUCOSE CONTROL SOLUTION LEVEL 2 SET

glucose dehydrogenase, glucose

Product Information

Product Type	MEDICAL DEVICE	Item Code (Source)	NHRIC:47884-181
---------------------	----------------	---------------------------	-----------------

Product Characteristics

number of times usable	4000
-------------------------------	------

Packaging

#	Item Code	Package Description	Marketing Start Date	Marketing End Date
1	NHRIC:47884-181-50	50 in 1 PACKAGE		

Marketing Information

Marketing Category	Application Number or Monograph Citation	Marketing Start Date	Marketing End Date
premarket notification	K101037	10/01/2013	

Labeler - EPS Bio Technology Corp. (658620117)

