

**A1181-18 CONTINUOUS EPIDURAL 18G HUSTEAD - regional anesthesia kit
Smiths Medical ASD, Inc.**

APLICARE POVIDONE-IODINE SOLUTION (povidone-iodine solution) solution
[Aplicare, Inc.]

3/4 Ounce Povidone Iodine Packet

Povidone-iodine 10%

Antiseptic

Warnings

Do not use

- if allergic to iodine
- in the eyes

For external use only

Ask a doctor before use if injuries are

- deep or puncture wounds
- serious burns

Stop use and ask a doctor if

- redness, irritation, swelling or pain persists or increases
- infection occurs

Avoid pooling beneath patient

Keep out of reach of children. In case of accidental ingestion, seek professional assistance or consult a poison control center immediately.

Package Label Display Panel

NDC 52380-0001-3
DIN 02076144

TEAR HERE
POUR



3/4 FLUID OUNCE
POVIDONE-IODINE
SOLUTION

0.75 Fl. oz. (22.5 mL)

STERILE unless opened or damaged.

Drug Facts	
Active ingredient	Purpose
Povidone-iodine USP 10%	Antiseptic
Use antiseptic skin preparation	
Warnings	
Do not use if allergic to iodine	
For external use only	

Reorder No. L-3001

EXP 10-10

TEAR HERE
POUR

Drug Facts (continued)

Warnings

Ask a doctor before use if injuries are
■ deep or puncture wounds ■ serious burns

Stop use and ask a doctor if

■ redness, irritation, swelling or pain persists or increases
■ infection occurs

Do not use in eyes

Avoid pooling beneath patient

Keep out of reach of children. In case of accidental ingestion, seek professional assistance or consult a poison control center immediately.

Directions apply locally as needed

Other information

■ 1% titratable iodine ■ latex free
■ for hospital or professional use only

Inactive ingredients citric acid, disodium phosphate, nonoxynol-9, sodium hydroxide, water


Questions or comments?

☎ 1-800-760-3236 (Mon to Fri 8:30 AM-5:00 PM EST)




APLICARE, INC.
MERIDEN, CT 06450
U.S.A.
DOLLARD-des-ORMEAUX,
QUEBEC H9G 2S3
www.aplicare.com
0407

Package Label Display Panel



REF A1181-18
CONTINUOUS EPIDURAL
18G HUSTEAD



PROCEDURAL COMPONENTS:

- 1 Hustead Epidural Needle (18G x 3 1/2 in. X.T.W.)
- 1 21G Portex® Nylon Epidural Catheter with Stylet, Open End, No Eyes for use with 17G or 18G Needle
- 1 EpiFuse™ Catheter Connector
- 1 Catheter Insertion Device
- 1 Pulsator® L.O.R. Syringe (7ml, Luer Slip)
- 1 Fenestrated Drape
- 1 Epidural Catheter I.D. Label
- 3 Gauze Sponges
- 1 18G x 1 1/2 in. Needle
- 1 22G x 1 1/2 in. Needle
- 1 25G x 5/8 in. Needle
- 1 27G x 1 1/2 in. T.W. Needle
- 1 Medicine Cup, 50ml
- 1 Plastic Syringe (10ml, Luer Lock)
- 1 Plastic Syringe (5ml, Luer Lock)
- 1 Plastic Syringe (3ml, Luer Lock)

PREP COMPONENTS:

- 2 Towels
- 3 Sponge Applicators
- 1 Povidone-Iodine Solution, 3/4 oz.

DEVICE INFORMATION:
CATHETER CONNECTOR
 See enclosed Instructions For Use.


WARNING(S):

- A needle-stick with a contaminated needle may cause infectious disease.

PRECAUTIONS:

- **Use Aseptic technique.**
- To help prevent needle-stick injuries, needles should not be recapped or purposely bent. If excessive resistance is met during needle insertion, do not force the needle as damage may occur. To help avoid needle breakage, do not attempt to straighten a bent needle; discard it and complete the procedure with a replacement needle.
- After use, place sharps in a suitable sharps container. Dispose of contaminated product in a safe manner according to Centers for Disease Control and Prevention (USA) and Federal/State/Local regulations (EPA, OSHA) and health care facility guidelines or local equivalent.
- Do NOT Resterilize.
- Never withdraw the catheter back through the epidural needle as this may cause the catheter to kink or shear.
- To be used only by individuals familiar with continuous epidural procedures. For specific techniques and procedures, refer to standard textbooks.


The Smiths Medical and Portex design marks; Portex; EpiFuse; and Pulsator are trademarks of the Smiths Medical family of companies. The symbol © indicates the trademark is registered in the U.S. Patent and Trademark Office and certain other countries.



STERILE EO Rx ONLY

Smiths Medical ASD, Inc.
 Keene, NH 03431 USA
 Made in USA
www.smiths-medical.com

PTA1181-18 REV.002 11/10 © 2010 Smiths Medical family of companies. All rights reserved. **smiths medical**



(01) 0 0351688 01713 3

PORTEX
CONTINUOUS EPIDURAL
18G HUSTEAD

A1181-18
 Use by: 0000-00
 yyyz-mm
 L01 000000

A1181-18 CONTINUOUS EPIDURAL 18G HUSTEAD

regional anesthesia kit kit

Product Information			
Product Type	MEDICAL DEVICE	Item Code (Source)	NHRIC:51688-8822

Packaging				
#	Item Code	Package Description	Marketing Start Date	Marketing End Date
1	NHRIC:51688-8822-2	10 in 1 CASE		
1		1 in 1 PACKAGE, COMBINATION		

Quantity of Parts		
Part #	Package Quantity	Total Product Quantity
Part 1	1 PACKET	22.5 mL

Part 1 of 1

APLICARE POVIDONE-IODINE

povidone-iodine solution

Product Information	
Item Code (Source)	NDC:52380-0001

Route of Administration TOPICAL

Active Ingredient/Active Moiety

Ingredient Name	Basis of Strength	Strength
POVIDONE-IODINE (UNII: 85H0HZU99M) (IODINE - UNII:9679TC07X4)	POVIDONE-IODINE	0.10 mg in 1 mL

Inactive Ingredients

Ingredient Name	Strength
CITRIC ACID MONOHYDRATE (UNII: 2968PHW8QP)	
SODIUM PHOSPHATE, DIBASIC (UNII: GR686LBA74)	
SODIUM HYDROXIDE (UNII: 55X04QC32I)	
NONOXYNOL-9 (UNII: 48Q180SH9T)	
WATER (UNII: 059QF0K00R)	

Packaging

#	Item Code	Package Description	Marketing Start Date	Marketing End Date
1	NDC:52380-0001-3	22.5 mL in 1 PACKET		

Marketing Information

Marketing Category	Application Number or Monograph Citation	Marketing Start Date	Marketing End Date
OTC mono graph not final	part333A	03/01/1984	

Marketing Information

Marketing Category	Application Number or Monograph Citation	Marketing Start Date	Marketing End Date
premarket notification	K965017	12/14/2005	

Labeler - Smiths Medical ASD, Inc. (137835299)

Establishment

Name	Address	ID/FEI	Business Operations
Smiths Medical ASD, Inc.		137835299	relabel, manufacture

Establishment

Name	Address	ID/FEI	Business Operations
Aplicare, Inc.		107255002	manufacture