

OXYGEN- oxygen gas
Alternacare Home Health Services

Disclaimer: This drug has not been found by FDA to be safe and effective, and this labeling has not been approved by FDA. For further information about unapproved drugs, click here.

WARNING:
HIGH PRESSURE GAS CYLINDER

Improper filling or abuse of this cylinder, or failure to heed this warning may cause serious injury or loss of life. DO NOT ALTER OR MODIFY CYLINDER OR RELATED COMPONENTS. DO NOT OVER PRESSURIZE. Keep out of the reach of children. Cylinders that have been refinished and/or exposed to elevated temperatures must be hydrostatically tested before filling. Do not use caustic paint strippers. Aluminum cylinders subject to action of fire or heated to temperatures in excess of 350 F must be withdrawn from service and condemned. Valves and safety relief devices must be removed and replaced by trained personnel. Do not alter or change safety discs. Replace pressure relief devices only with complete assemblies supplied by the valve manufacturer. DO NOT FILL WITH OXIDIZING GASES WITHOUT PROPER CLEANING. Cylinders, valves and assemblies must be specially cleaned for oxygen. Verify cylinder content and pressure monthly, or more frequently if required. Store and use with adequate ventilation.

WARNING: For emergency use only when administered by properly trained personnel for oxygen deficiency and resuscitation. For all other medical applications **Rx ONLY**. Uninterrupted use of high concentrations of oxygen over a long duration, without monitoring its effect on oxygen content of arterial blood, may be harmful. Do not attempt to use on patients who have stopped breathing unless used in conjunction with resuscitative equipment.

WARNING: HIGH PRESSURE OXIDIZING GAS VIGOROUSLY ACCELERATED COMBUSTION

Use only with pressure reducing equipment and apparatus cleaned and designed for oxygen service and rated for cylinder pressure.

Open valve slowly. Close valve after each use and when empty. Keep cylinder, cylinder valve, all equipment that connects, free of oil and grease to avoid violent ignition.

Federal law requires that this container be refilled with oxygen USP only by establishments registered as a drug manufacturer in accordance with the Federal Food, Drug, and Cosmetic Act.

Pursuant to recommendations by Compressed Gas Association, do not expose pressurized cylinders to temperatures in excess of 125 F

DO NOT REMOVE THIS PRODUCT LABEL

Use a back flow preventative device in the piping.

Use only in accordance with product instructions and material safety data sheet.

OXYGEN

COMPRESSED

U.S.P. UN 1072

Produced by Air Liquefaction

OXYGEN - 2

Cylinder Contents at 70 F

Type B - 142 liters at 1900 psig

Type C - 235 liters at 1900 psig

Type D - 393 liters at 1900 psig

Type E - 646 liters at 1900 psig

Date code 1104 Label 0035-17(1)

Produced by air liquefaction

OXYGEN COMPRESSED U.S.P. UN 1072



OXYGEN
2

Distributed by:

WARNING:
HIGH PRESSURE GAS CYLINDER
Improper filling or abuse of this cylinder, or failure to heed this warning may cause serious injury or loss of life. DO NOT ALTER OR MODIFY CYLINDER OR RELATED COMPONENTS. DO NOT OVER PRESSURIZE. Keep out of the reach of children. Cylinders that have been refilled and/or exposed to elevated temperatures must be hydrostatically tested before filling. Do not use caustic paint strippers. Aluminum cylinders subject to action of fire or heated to temperatures in excess of 350°F must be withdrawn from service and condemned. Valves and safety relief devices must be removed and replaced by trained personnel. Do not alter or change safety devices. Replace pressure relief devices only with complete assemblies supplied by the valve manufacturer. DO NOT FILL WITH OXIDIZING GASES WITHOUT PROPER CLEANING. Cylinders, valves, and assemblies must be specially cleaned for oxygen. Verify cylinder content and pressure monthly, or more frequently if required. Store and use with adequate ventilation.

WARNING: For emergency use only when administered by properly trained personnel for oxygen deficiency and resuscitation. For all other medical applications: Its ONLY. Uninterrupted use of high concentrations of oxygen over a long duration, without monitoring its effect on oxygen content of arterial blood, may be harmful. Do not attempt to use on patients who have stopped breathing unless used in conjunction with resuscitative equipment.

WARNING: HIGH PRESSURE OXIDIZING GAS
VIGOROUSLY ACCELERATES COMBUSTION

- Use only with pressure reducing equipment and apparatus cleaned and designed for oxygen service and rated for cylinder pressure.
- Open valve slowly. Close valve after each use and when empty. Keep cylinder, cylinder valve, all equipment that connects, free of oil and grease to avoid violent ignition.
- Federal law requires that this container be refilled with oxygen USP only by establishments registered as a drug manufacturer in accordance with the Federal Food, Drug, and Cosmetic Act.
- Pursuant to recommendations by Compressed Gas Association, do not expose pressurized cylinders to temperatures in excess of 125°F.
- DO NOT REMOVE THIS PRODUCT LABEL.
Use a back flow preventive device in the piping.
Use only in accordance with product instructions and material safety data sheet.

Cylinder contents at 70°F

Type	Contents
<input type="radio"/> B	142 liters at 1900 PSI
<input type="radio"/> C	235 liters at 1900 PSI
<input type="radio"/> D	393 liters at 1900 PSI
<input type="radio"/> E	646 liters at 1900 PSI

res

OXYGEN				
oxygen gas				
Product Information				
Product Type	HUMAN PRESCRIPTION DRUG	Item Code (Source)	NDC:64097-0001	
Route of Administration	RESPIRATORY (INHALATION)			
Active Ingredient/Active Moiety				
	Ingredient Name	Basis of Strength	Strength	
	OXYGEN (UNII: S88TT14065) (OXYGEN - UNII:S88TT14065)	OXYGEN	99.5 L in 100 L	
Packaging				
#	Item Code	Package Description	Marketing Start Date	Marketing End Date

1	NDC:64097-0001-1	142 L in 1 CYLINDER; Type 0: Not a Combination Product	01/01/1996	
2	NDC:64097-0001-2	235 L in 1 CYLINDER; Type 0: Not a Combination Product	01/01/1996	
3	NDC:64097-0001-3	393 L in 1 CYLINDER; Type 0: Not a Combination Product	01/01/1996	
4	NDC:64097-0001-4	646 L in 1 CYLINDER; Type 0: Not a Combination Product	01/01/1996	

Marketing Information

Marketing Category	Application Number or Monograph Citation	Marketing Start Date	Marketing End Date
unapproved medical gas		01/01/1996	

Labeler - Alternacare Home Health Services (946763364)

Registrant - Alternacare Home Health Services (946763364)

Establishment

Name	Address	ID/FEI	Business Operations
Alternacare Home Health Services		946763364	manufacture(64097-0001)

Revised: 10/2023

Alternacare Home Health Services