

OXYGEN- oxygen gas
Praxair Distribution Mid-atlantic, Llc

Oxygen

UN1072 OXYGEN, COMPRESSED USP WARNING! HIGH-PRESSURE, OXIDIZING GAS. VIGOROUSLY ACCELERATES COMBUSTION. Keep oil, grease and combustibles away. Keep away from heat, sparks and flame. Store and use with adequate ventilation to prevent oxygen accumulation. Use with equipment rated for cylinder pressure and cleaned for oxygen service. Use a backflow prevention device in any piping. Cylinder temperature should not exceed 125 deg F (52 deg C). Open valve slowly. Close valve after each use; keep closed even when empty. Always secure cylinder. Install cap, if provided, when not in use. Use per Praxair MSDS form P-4638 and safety booklets P-3499, P-14-153, and P-15-276; obtain from your local supplier. KEEP OUT OF REACH OF CHILDREN. WARNING! For emergency use only when administered by properly trained personnel for oxygen deficiency and resuscitation. For all other medical applications, Rx ONLY. Uninterrupted use of high concentrations of oxygen over a long duration, without monitoring its effects on oxygen content of arterial blood, may be harmful. Use only with pressure reducing equipment and apparatus designed for oxygen. Do not attempt to use on patients who have stopped breathing unless used in conjunction with resuscitative equipment. Rx ONLY CAS 7782-44-7 PRODUCED BY AIR LIQUEFACTION IN CASE OF EMERGENCY: CALL 1-800-645-4633 DISTRIBUTED BY PRAXAIR, INC., DANBURY DO NOT REMOVE THIS LABEL PMG-6 (07/08) Praxair, the flowing airstream design and Medipure are trademarks of Praxair Technology, Inc. DISTRIBUTED BY PRAXAIR, INC, DANBURY, CT 06810-5113 DISTRIBUTED BY PRAXAIR, INC., DANBURY



UN1072 OXYGEN, COMPRESSED USP DANGER! COMPRESSED, OXIDIZING GAS. MAY CAUSE OR INTENSIFY FIRE; OXIDIZER. CONTAINS GAS UNDER PRESSURE; MAY EXPLODE IF HEATED. Keep/store away from combustible materials. Keep reduction valves free from grease and oil. In case of fire: stop leak if safe to do so. Open valve slowly. Close valve after each use and when empty. Use only with equipment cleaned for oxygen service and rated for cylinder pressure. Use a backflow preventive device in the piping. Protect from sunlight when ambient temperature exceeds 52 deg C/125 deg F. Cylinder temperature should not exceed 52 deg C/125 deg F. Store and use in a well-ventilated place. Use per Praxair MSDS form P-4638, and safety booklets P-3499, P-14-153, and P-15-276; obtain from your local supplier. KEEP OUT OF REACH OF CHILDREN. Do not handle until all safety precautions have been read and understood. WARNING! For emergency use only when administered by properly trained personnel for oxygen deficiency and resuscitation. For all other medical applications, Rx ONLY. Uninterrupted use of high concentrations of oxygen over a long duration, without monitoring its effects on oxygen content of arterial blood, may be harmful. Use only

with pressure reducing equipment and apparatus designed for oxygen. Do not attempt to use on patients who have stopped breathing unless used in conjunction with resuscitative equipment DO NOT REMOVE THIS LABEL Rx ONLY CAS 7782-44-7 PRODUCED BY AIR LIQUEFACTION IN CASE OF EMERGENCY: CALL 1-800-645-4633 DISTRIBUTED BY PRAXAIR, INC., DANBURY PMGB-6 (03/11) Praxair, the flowing airstream design and Medipure are trademarks of Praxair Technology, Inc. DISTRIBUTED BY PRAXAIR, INC, DANBURY, CT 06810-5113

**UN1072
OXYGEN,
COMPRESSED
USP**



DANGER! COMPRESSED, OXIDIZING GAS. MAY CAUSE OR INTENSIFY FIRE; OXIDIZER. CONTAINS GAS UNDER PRESSURE; MAY EXPLODE IF HEATED. Keep/store away from combustible materials. Keep reduction valves free from grease and oil. In case of fire: stop leak if safe to do so. Open valve slowly. Close valve after each use and when empty. Use only with equipment cleaned for oxygen service and rated for cylinder pressure. Use a backflow preventive device in the piping. Protect from sunlight when ambient temperature exceeds 52°C/125°F. Cylinder temperature should not exceed 52°C/125°F. Store and use in a well-ventilated place. Use per Praxair MSDS form P-4638, and safety booklets P-3499, P-14-153, and P-15-276; obtain from your local supplier. **KEEP OUT OF REACH OF CHILDREN.** Do not handle until all safety precautions have been read and understood.

WARNING! For emergency use only when administered by properly trained personnel for oxygen deficiency and resuscitation. For all other medical applications, **Rx ONLY.** Uninterrupted use of high concentrations of oxygen over a long duration, without monitoring its effects on oxygen content of arterial blood, may be harmful. Use only with pressure reducing equipment and apparatus designed for oxygen. Do not attempt to use on patients who have stopped breathing unless used in conjunction with resuscitative equipment.

DO NOT REMOVE THIS LABEL

IN CASE OF EMERGENCY: CALL 1-800-645-4633
PRODUCED BY AIR LIQUEFACTION CAS: 7782-44-7



PMGB-6 (03/11)

DISTRIBUTED BY PRAXAIR, INC.,
DANBURY, CT 06810-5113

Praxair, the Flowing Airstream design, and Medipure are trademarks of Praxair Technology, Inc.

UN1073 OXYGEN, REFRIGERATED LIQUID USP ALWAYS KEEP CONTAINER IN UPRIGHT POSITION. DO NOT CHANGE OR FORCE-FIT CONNECTIONS LOT # NET CONTENTS WARNING! EXTREMELY COLD, OXIDIZING LIQUID AND GAS UNDER PRESSURE. VIGOROUSLY ACCELERATES COMBUSTION. COMBUSTIBLES IN CONTACT WITH LIQUID OXYGEN MAY EXPLODE ON IGNITION OR IMPACT. CAN CAUSE SEVERE FROSTBITE.

MAY CAUSE DIZZINESS OR DROWSINESS. ODOR: NONE Do not drop. Use suitable hand truck for container movement. Keep oil, grease, and combustibles away. Keep away from heat, sparks, and flame. Use with equipment cleaned for oxygen service. Do not get liquid in eyes, on skin, or clothing. For liquid withdrawal, wear face shield and gloves. Avoid spills. Do not walk on or roll equipment over spills. Store and use with adequate ventilation to prevent oxygen accumulation. Container temperature should not exceed 125 deg F (52 deg C). Use piping and equipment adequately designed to withstand any pressures to be encountered. Use a backflow prevention device in any piping. Close valve after each use; keep closed even when empty. Container discharge must be warmed to room temperature before breathing. Always secure container so it cannot be knocked over. Use per Praxair MSDS form P-4637 and safety booklets P-3499, P-14-153, and P-15-276 (obtain from your local supplier), and per manufacturer's instructions for this container. RX ONLY FIRST AID: INHALATION—immediately remove to fresh air. If not breathing, give artificial respiration. Keep victim warm and at rest. Call a physician. Advise physician that the victim has been exposed to a high concentration of oxygen. FROSTBITE—Obtain medical treatment immediately. FOR SKIN CONTACT, immediately warm frostbite area with warm water not exceeding 105 deg F (41 deg C). In case of massive exposure, remove clothing while showering with warm water. Call a physician. FOR EYE CONTACT, flush eyes with warm water for at least 15 minutes. See a physician, preferably an ophthalmologist, immediately. WARNING: For emergency use only when administered by properly trained personnel for oxygen deficiency and resuscitation. For all other medical applications, Rx ONLY. Uninterrupted use of high concentrations of oxygen over a long duration, without monitoring its effects on oxygen content of arterial blood, may be harmful. Use only with pressure reducing equipment and apparatus designed for oxygen. Do not attempt to use on patients who have stopped breathing unless used in conjunction with resuscitative equipment. DO NOT REMOVE THIS LABEL Rx ONLY CAS 7782-44-7 PRODUCED BY AIR LIQUEFACTION IN CASE OF EMERGENCY: CALL 1-800-645-4633 PMG-8 (07/08) Copyright @ 2002-2003, 2008. Praxair Technology, Inc. Praxair, the flowing airstream design and Medipure are trademarks of Praxair Technology, Inc. DISTRIBUTED BY PRAXAIR, INC, DANBURY, CT 06810-5113

**UN 1073
OXYGEN,
REFRIGERATED LIQUID
USP**

CAS: 7782-44-7
PRODUCED BY AIR LIQUEFACTION

WARNING! EXTREMELY COLD, OXIDIZING LIQUID AND GAS UNDER PRESSURE. VIGOROUSLY ACCELERATES COMBUSTION. COMBUSTIBLES IN CONTACT WITH LIQUID OXYGEN MAY EXPLODE ON IGNITION OR IMPACT. CAN CAUSE SEVERE FROSTBITE. MAY CAUSE DIZZINESS OR DROWSINESS. ODOR: NONE

Do not drop. Use suitable hand truck for container movement. Keep oil, grease, and combustibles away. Keep away from heat, sparks, and flame. Use with equipment cleaned for oxygen service. Do not get liquid in eyes, on skin, or clothing. For liquid withdrawal, wear face shield and gloves. Avoid spills. Do not walk on or roll equipment over spills. Store and use with adequate ventilation to prevent oxygen accumulation. Container temperature should not exceed 125°F (52°C). Use piping and equipment adequately designed to withstand any pressures to be encountered. Use a backflow prevention device in any piping. Close valve after each use; keep closed even when empty. Container discharge must be warmed to room temperature before breathing. Always secure container so it cannot be knocked over. Use per Praxair MSDS form P-4637 and safety booklets P-3499, P-14-153, and P-15-276 (obtain from your local supplier), and per manufacturer's instructions for this container.

IN CASE OF EMERGENCY: CALL 1-800-645-4633

MADE IN USA
PMG-8 (07/08)



OXYGEN

2



↑ ALWAYS KEEP CONTAINER IN UPRIGHT POSITION ↑
DO NOT CHANGE OR FORCE-FIT CONNECTIONS

LOT NUMBER _____

NET CONTENTS _____

Rx ONLY

FIRST AID: INHALATION — Immediately remove to fresh air. If not breathing, give artificial respiration. Keep victim warm and at rest. Call a physician. Advise physician that the victim has been exposed to a high concentration of oxygen. FROSTBITE—Obtain medical treatment immediately. FOR SKIN CONTACT, immediately warm frostbite area with warm water not exceeding 105°F (41°C). In case of massive exposure, remove clothing while showering with warm water. Call a physician. FOR EYE CONTACT, flush eyes with warm water for at least 15 minutes. See a physician, preferably an ophthalmologist, immediately.

WARNING: For emergency use only when administered by properly trained personnel for oxygen deficiency and resuscitation. For all other medical applications, Rx ONLY. Uninterrupted use of high concentrations of oxygen over a long duration, without monitoring its effects on oxygen content of arterial blood, may be harmful. Use only with pressure reducing equipment and apparatus designed for oxygen. Do not attempt to use on patients who have stopped breathing unless used in conjunction with resuscitative equipment.

DO NOT REMOVE THIS LABEL



DISTRIBUTED BY PRAXAIR, INC.
DANBURY, CT 06810-5113

Copyright © 2002-2003, 2008, Praxair Technology, Inc.
Praxair, the Flowing Airstream design, and Medipure are trademarks of Praxair Technology, Inc.

OXYGEN

oxygen gas

Product Information

Product Type	HUMAN PRESCRIPTION DRUG	Item Code (Source)	NDC:76 133-001
Route of Administration	RESPIRATORY (INHALATION)		

Active Ingredient/Active Moiety

Ingredient Name	Basis of Strength	Strength
OXYGEN (UNII: S88TT14065) (OXYGEN - UNII:S88TT14065)	OXYGEN	990 mL in 1 L

Packaging

#	Item Code	Package Description	Marketing Start Date	Marketing End Date
1	NDC:76 133-001-01	16424 L in 1 CYLINDER; Type 0: Not a Combination Product		
2	NDC:76 133-001-02	14130 L in 1 CYLINDER; Type 0: Not a Combination Product		
3	NDC:76 133-001-03	10987 L in 1 CYLINDER; Type 0: Not a Combination Product		
4	NDC:76 133-001-04	9543 L in 1 CYLINDER; Type 0: Not a Combination Product		
5	NDC:76 133-001-05	7985 L in 1 CYLINDER; Type 0: Not a Combination Product		
6	NDC:76 133-001-06	7051 L in 1 CYLINDER; Type 0: Not a Combination Product		
7	NDC:76 133-001-07	4361 L in 1 CYLINDER; Type 0: Not a Combination Product		
8	NDC:76 133-001-08	3511 L in 1 CYLINDER; Type 0: Not a Combination Product		
9	NDC:76 133-001-09	2350 L in 1 CYLINDER; Type 0: Not a Combination Product		
10	NDC:76 133-001-10	708 L in 1 CYLINDER; Type 0: Not a Combination Product		
11	NDC:76 133-001-11	595 L in 1 CYLINDER; Type 0: Not a Combination Product		
12	NDC:76 133-001-12	425 L in 1 CYLINDER; Type 0: Not a Combination Product		
13	NDC:76 133-001-13	142 L in 1 CYLINDER; Type 0: Not a Combination Product		
14	NDC:76 133-001-14	198 L in 1 CYLINDER; Type 0: Not a Combination Product		
15	NDC:76 133-001-15	7730 L in 1 CYLINDER; Type 0: Not a Combination Product		
16	NDC:76 133-001-16	4332 L in 1 CYLINDER; Type 0: Not a Combination Product		
17	NDC:76 133-001-17	2520 L in 1 CYLINDER; Type 0: Not a Combination Product		
18	NDC:76 133-001-18	934 L in 1 CYLINDER; Type 0: Not a Combination Product		
19	NDC:76 133-001-19	680 L in 1 CYLINDER; Type 0: Not a Combination Product		
20	NDC:76 133-001-20	142 L in 1 CYLINDER; Type 0: Not a Combination Product		
21	NDC:76 133-001-21	74566 L in 1 DEWAR; Type 0: Not a Combination Product		
22	NDC:76 133-001-22	103757 L in 1 DEWAR; Type 0: Not a Combination Product		
23	NDC:76 133-001-23	114228 L in 1 DEWAR; Type 0: Not a Combination Product		
24	NDC:76 133-001-24	120574 L in 1 DEWAR; Type 0: Not a Combination Product		
25	NDC:76 133-001-25	127237 L in 1 DEWAR; Type 0: Not a Combination Product		
26	NDC:76 133-001-26	135487 L in 1 DEWAR; Type 0: Not a Combination Product		
27	NDC:76 133-001-27	142785 L in 1 DEWAR; Type 0: Not a Combination Product		

28	NDC:76133-001-28	155794 L in 1 DEWAR; Type 0: Not a Combination Product		
29	NDC:76133-001-29	162140 L in 1 DEWAR; Type 0: Not a Combination Product		
30	NDC:76133-001-30	178957 L in 1 DEWAR; Type 0: Not a Combination Product		
31	NDC:76133-001-31	291916 L in 1 DEWAR; Type 0: Not a Combination Product		
32	NDC:76133-001-32	428355 L in 1 DEWAR; Type 0: Not a Combination Product		

Marketing Information

Marketing Category	Application Number or Monograph Citation	Marketing Start Date	Marketing End Date
NDA	NDA205849	10/01/2007	

Labeler - Praxair Distribution Mid-atlantic, Llc (808461896)

Establishment

Name	Address	ID/FEI	Business Operations
Praxair Distribution Mid-atlantic, Llc		808461896	manufacture(76133-001)

Establishment

Name	Address	ID/FEI	Business Operations
Praxair Distribution, Inc.		001355176	manufacture(76133-001)

Establishment

Name	Address	ID/FEI	Business Operations
Praxair Distribution, Inc.		623107120	manufacture(76133-001)

Establishment

Name	Address	ID/FEI	Business Operations
Praxair Distribution, Inc.		167418326	manufacture(76133-001)

Establishment

Name	Address	ID/FEI	Business Operations
Praxair Distribution, Inc.		063205512	manufacture(76133-001)

Establishment

Name	Address	ID/FEI	Business Operations
Praxair Distribution, Inc.		039755707	manufacture(76133-001)

Establishment

Name	Address	ID/FEI	Business Operations
Praxair Distribution, Inc.		002246916	manufacture(76133-001)

Establishment

Name	Address	ID/FEI	Business Operations

Praxair Distribution, Inc.		364098731	manufacture(76133-001)
----------------------------	--	-----------	------------------------

Establishment

Name	Address	ID/FEI	Business Operations
Praxair Distribution, Inc.		824679554	manufacture(76133-001)

Establishment

Name	Address	ID/FEI	Business Operations
Praxair Distribution, Inc.		148052277	manufacture(76133-001)

Establishment

Name	Address	ID/FEI	Business Operations
Praxair Distribution, Inc.		879317022	manufacture(76133-001)

Establishment

Name	Address	ID/FEI	Business Operations
Praxair Distribution, Inc.		021450408	manufacture(76133-001)

Establishment

Name	Address	ID/FEI	Business Operations
Praxair Distribution, Inc.		555510726	manufacture(76133-001)

Establishment

Name	Address	ID/FEI	Business Operations
Praxair Distribution, Inc.		037432010	manufacture(76133-001)

Establishment

Name	Address	ID/FEI	Business Operations
Praxair Distribution, Inc.		099430951	manufacture(76133-001)

Establishment

Name	Address	ID/FEI	Business Operations
Praxair Distribution, Inc.		884877440	manufacture(76133-001)

Establishment

Name	Address	ID/FEI	Business Operations
Praxair Distribution, Inc.		118788132	manufacture(76133-001)

Establishment

Name	Address	ID/FEI	Business Operations
Praxair Distribution, Inc.		623107070	manufacture(76133-001)

Establishment

Name	Address	ID/FEI	Business Operations
Praxair Distribution, Inc.		623107088	manufacture(76133-001)

